

## ROMED HOLLAND Declaration of Conformity

<b>Manufacturer</b>	Van Oostveen Medical BV Herenweg 269 3648 CH Wilnis Netherlands	<b>Tel: 31-297-282 101</b> <b>Fax: 31-297-288316</b> <b>E-mail: <a href="mailto:info@romed.nl">info@romed.nl</a></b> <b>Website : <a href="http://www.romed.nl">www.romed.nl</a></b>
<b>SRN</b>	<b>NL-MF-000003481</b>	
<b>Product</b>	Plaster of Paris Ref: POP050-300, POP075-300, POP100-300, POP150-300, POP200-300, POP075-270, POP100-270, POP150-270, POP200-270	
<b>Brand</b>	ROMED	
<b>Classification (MDR, Annex VIII)</b>	Class I	
<b>UMDNS</b>	11327	
<b>GMDN</b>	33056	
<b>Basic UDI-DI</b>	871720243MDIPOP000001ML	

We, with sole responsibility in drawing up this Declaration of Conformity, declare that the above mentioned product meets the provisions of the following EC Council Directives, Regulations and Standards. All supporting documentation is retained under the premises of the manufacturer.

### General applicable directives and regulations

REGULATION (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices

### Applicable standards

EN ISO 13485:2016, EN ISO 14971:2019, EN ISO 15223-1:2016, EN 1041:2008+A1:2013.

No EU common specifications are applicable.

Wilnis, The Netherlands, 21 May 2021

Signed on behalf of Van Oostveen Medical BV



M.J. van Oostveen  
Managing Director