## **VAN OOSTVEEN** MEDICAL B.V.

MEDICAL DISPOSABLES



Herenweg 269 3648 CH Wilnis The Netherlands Tel. +31 297 28 21 01 Fax +31 297 28 83 16 E-mail info@romed.nl www.romed.nl

## ROMED HOLLAND Declaration of Conformity

Manufacturer	Van Oostveen Medical BV	Tel: 31-297-282 101
	Herenweg 269	Fax: 31-297-288316
	3648 CH Wilnis	E-mail: info@romed.nl
	Netherlands	Website : www.romed.nl
SRN	NL-MF-000003481	
Product	Lemon Glycerine Swabsticks	
	Ref: LSWAB-250	
Brand	ROMED	
Classification (MDR, Annex VIII)	Class I	
UMDNS	16619	
GMDN	48037	
Basic UDI-DI	871720243MDISOCDN000001F6	

We, with sole responsibility in drawing up this Declaration of Conformity, declare that the above mentioned product meets the provisions of the following EC Council Directives, Regulations and Standards. All supporting documentation is retained under the premises of the manufacturer.

## General applicable directives and regulations

REGULATION (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices

## Applicable standards

EN ISO 13485:2016, EN ISO 14971:2019, EN ISO 15223-1:2016, EN 1041:2008+A1:2013. No EU common specifications are applicable.

Wilnis, The Netherlands, 21 May 2021 Signed on behalf of Van Oostveen Medical BV



M.J. van Oostveen Managing Director